

**DEPARTAMENTO DE MANTENIMIENTO**

REF R-M-13
FECHA 11-04-2019

REGISTRO CALIBRACION DE CONDUCTIVIDAD CT-903

FECHA	LECTURA INICIAL (uS)	LECTURA LABORATORIO (uS)	LECTURA FINAL (uS)	OBSERVACIONES
05-09-2016	3353	3490	3490	CALIBRACION OK
03-10-2016	3111	3215	3215	CALIBRACION OK
02-11-2016	2944	3080	3080	CALIBRACION OK
05-12-2016	3120	3213	3213	CALIBRACION OK
03-01-2017	3219	3250	3250	CALIBRACION OK
19-01-2017	2011	2190	2120	LIMPIAR SENSOR
13-02-2017	3224	3210	3210	CALIBRACION OK
13-03-2017	3315	3300	3300	CALIBRACION OK
12-04-2017	3115	3050	3050	CALIBRACION OK
10-05-2017	2640	2740	2740	CALIBRACION OK
09-06-2017	3150	3210	3210	CALIBRACION OK
10-07-2017	2780	2950	2950	CALIBRACION OK
10-08-2017	2850	3000	3000	CALIBRACION OK
11-09-2017	2910	3020	3020	CALIBRACION OK
11-10-2017	2800	2890	2890	CALIBRACION OK
10-11-2017	3023	3050	3050	CALIBRACION OK
11-12-2017	2950	3000	3000	CALIBRACION OK
11-01-2018	3100	3130	3130	CALIBRACION OK
09-02-2018	2920	2960	2960	CALIBRACION OK
09-03-2018	3120	3150	3150	CALIBRACION OK
09-04-2018	2830	2880	2880	CALIBRACION OK
09-05-2018	2960	2970	2970	CALIBRACION OK
11-06-2018	3310	3325	3325	CALIBRACION OK
11-07-2018	3030	3040	3040	CALIBRACION OK
09-08-2018	2940	2960	2960	CALIBRACION OK
10-09-2018	3053	3100	3100	CALIBRACION OK
10-10-2018	2832	2850	2850	CALIBRACION OK
12-11-2018	3122	3150	3150	CALIBRACION OK
12-12-2018	3026	3050	3050	CALIBRACION OK
11-01-2019	2846	2855	2855	CALIBRACION OK
11-02-2019	2946	3000	3000	CALIBRACION OK
11-03-2019	3125	3150	3150	CALIBRACION OK
11-04-2019	3150	3100	3100	CALIBRACION OK