

**DEPARTAMENTO DE MANTENIMIENTO**

REF R-M-13
FECHA 11/04/2019

REGISTRO CALIBRACION DE CONDUCTIVIDAD CT-903

FECHA	LECTURA INICIAL (uS)	LECTURA LABORATORIO (uS)	LECTURA FINAL (uS)	OBSERVACIONES
05/09/2016	3353	3490	3490	CALIBRACION OK
03/10/2016	3111	3215	3215	CALIBRACION OK
02/11/2016	2944	3080	3080	CALIBRACION OK
05/12/2016	3120	3213	3213	CALIBRACION OK
03/01/2017	3219	3250	3250	CALIBRACION OK
19/01/2017	2011	2190	2120	LIMPIAR SENSOR
13/02/2017	3224	3210	3210	CALIBRACION OK
13/03/2017	3315	3300	3300	CALIBRACION OK
12/04/2017	3115	3050	3050	CALIBRACION OK
10/05/2017	2640	2740	2740	CALIBRACION OK
09/06/2017	3150	3210	3210	CALIBRACION OK
10/07/2017	2780	2950	2950	CALIBRACION OK
10/08/2017	2850	3000	3000	CALIBRACION OK
11/09/2017	2910	3020	3020	CALIBRACION OK
11/10/2017	2800	2890	2890	CALIBRACION OK
10/11/2017	3023	3050	3050	CALIBRACION OK
11/12/2017	2950	3000	3000	CALIBRACION OK
11/01/2018	3100	3130	3130	CALIBRACION OK
09/02/2018	2920	2960	2960	CALIBRACION OK
09/03/2018	3120	3150	3150	CALIBRACION OK
09/04/2018	2830	2880	2880	CALIBRACION OK
09/05/2018	2960	2970	2970	CALIBRACION OK
11/06/2018	3310	3325	3325	CALIBRACION OK
11/07/2018	3030	3040	3040	CALIBRACION OK
09/08/2018	2940	2960	2960	CALIBRACION OK
10/09/2018	3053	3100	3100	CALIBRACION OK
10/10/2018	2832	2850	2850	CALIBRACION OK
12/11/2018	3122	3150	3150	CALIBRACION OK
11/01/2019	2846	2855	2855	CALIBRACION OK
11/02/2019	2946	3000	3000	CALIBRACION OK
11/03/2019	3125	3150	3150	CALIBRACION OK
11/04/2019	3150	3100	3100	CALIBRACION OK