



October 31, 2016

CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

FLUID	VISCOSITY	TEMPERATURE	LOT NUMBER
50	48.4	25.0	103116

	(CENTIPOISE)	(°C)
MODEL	TEST INSTRUMENT	
LVDV-II+		SERIAL NUMBER
		6524090
MODEL	TEMPERATURE SENSOR	
		SERIAL NUMBER
1529		A56933

CALIBRATION OF THE ABOVE MENTIONED BROOKFIELD VISCOSITY STANDARD IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.
THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH A BROOKFIELD VISCOMETER
THAT HAS BEEN CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE
OF STANDARDS & TECHNOLOGY.

BROOKFIELD ENGINEERING LABORATORIES, INC.

11 Commerce Blvd., Middleboro, MA 02346-1031 U.S.A.

TEL 800-628-8139 or 508-946-6200 FAX 508-946-6262

www.brookfieldengineering.com

BROOKFIELD ENGINEERING LABORATORIES, INC.

A handwritten signature in blue ink that reads "D.J. Moonay".

DAVID J. MOONAY, Ph.D. - RHEOLOGY LABORATORY SUPERVISOR