



November 29, 2016

CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

FLUID	VISCOSITY	TEMPERATURE	LOT NUMBER
500	498	25.0	112916

(CENTIPOISE)

(°C)

MODEL

TEST INSTRUMENT

SERIAL NUMBER

LVDV-II+

TEMPERATURE SENSOR

6529213

MODEL

SERIAL NUMBER

1529

A56933

CALIBRATION OF THE ABOVE MENTIONED BROOKFIELD VISCOSITY STANDARD IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.
THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH A BROOKFIELD VISCOMETER
THAT HAS BEEN CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE
OF STANDARDS & TECHNOLOGY.

BROOKFIELD ENGINEERING LABORATORIES, INC.

11 Commerce Blvd., Middleboro, MA 02346-1031 U.S.A.

TEL 800-628-8139 or 508-946-6200 FAX 508-946-6262

www.brookfieldengineering.com

BROOKFIELD ENGINEERING LABORATORIES, INC.

A handwritten signature in blue ink that reads "D.J. Mooney".

DAVID J. MOONAY, Ph.D - RHEOLOGY LABORATORY SUPERVISOR