

	REGISTRO CALIBRACION PEACHIMETRO	Código	
		Revisión	000
		Fecha	31-05-2016
		Página	1 de 1
		Revisado por	E. Pizarro
		Aprobado por	A. Greig

Equipo:	Orion Star A 2011			Código:	PML-017		Turno	
Fecha	Buffer Ph 7,00 +/- 0,05	Buffer Ph 4,00 +/- 0,05	Buffer Ph 10,00 +/- 0,05	Pendiente %	Temperatura °C	Responsable	Día	Noche
21-02-2017	7.00	4.01	10.02	97.3%	23.8	Jarol D		✓
21-02-2017	7.02	4.00	10.00	98.4%	18.9	ESTebanc	✓	
22-02-2017	7.00	4.01	10.03	97.5%	23.2	Jarol D		✓
22-02-2017	7.01	4.00	10.00	97.7%	20.0	ESTebanc	✓	
23-02-17	7.01	4.01	10.00	98.0	20.5	Jarol D		✓
23-02-17	7.02	4.00	10.00	97.9%	18.4	ESTebanc	✓	
24-02-17	7.01	4.00	10.00	97.7	18.5	Jarol D		✓
24-02-17	7.01	4.01	10.00	98.0%	20.4	ESTebanc	✓	
25-02-17	7.01	4.00	10.00	97.5	20.5	Jarol D		✓
25-02-17	7.01	4.00	10.00	97.4	19.6	ESTebanc	✓	
26-02-17	7.01	4.00	10.00	96.3	19.4	Jarol D		✓
26-02-17	7.01	4.01	10.01	97.0%	20.8	Jarol D	✓	
27-02-17	7.00	4.01	10.00	96.4	22.5	Jarol D		✓
27-02-17	7.01	4.01	10.01	97.0%	20.8	Jarol D	✓	
28-02-17	7.01	4.01	10.04	96.0%	22.5	Jarol D		✓
28-02-17	7.01	4.00	10.01	96.7%	19.9	Jarol D	✓	
01-03-17	7.01	4.01	10.04	95.9%	22.1	ESTebanc		✓
1-03-17	7.02	4.00	10.00	96.8	18.7	Conle E		✓
02-03-17	7.00	4.01	10.04	95.6%	22.4	ESTebanc		✓
02-03-17	7.02	4.00	10.00	96.1	19.3	Conle E		✓
03-03-17	7.01	4.01	10.05	95.2%	21.4	ESTebanc		✓
03-03-17	7.02	4.00	10.00	96.1	18.1	Conle E		✓
04-03-17	7.01	4.01	10.00	96.0	20.5	ESTebanc		✓
04-03-17	7.01	4.00	10.00	95.8	20.2	Jarol D	✓	
05-03-17	7.01	4.01	10.04	94.9	22.0	ESTebanc		✓
05-03-17	7.01	4.00	10.00	95.4	19.6	Jarol D	✓	
06-03-17	7.01	4.01	10.05	94.4	21.0	ESTebanc		✓
06-03-17	7.02	4.00	10.00	95.1	18.5	Jarol D	✓	
06-03-17	7.00	4.01	10.02	94.3	23.2	Conle E		✓
07-03-17	7.02	4.00	10.00	94.9	18.1	Jarol D	✓	
08-03-17	7.01	4.02	10.04	93.8	22.0	Conle E		✓
08-03-17	7.01	4.01	10.00	94.7	20.5	Jarol D	✓	
08-03-17	7.01	4.01	10.00	94.5	21.7	Conle E		✓
09-03-17	7.02	4.00	10.00	94.6	17.8	Jarol D	✓	
09-03-17	7.01	4.00	10.00	93.7	21.4	Conle E		✓
10-03-17	7.02	4.00	10.00	94.6	17.5	ESTebanc	✓	
10-03-17	7.01	4.01	10.00	94.4	21.7	Conle E		✓