

	CONTROL DE CALIDAD	
	REGISTRO DE VERIFICACION FUNCIONAMIENTO CONDUCTIMETRO	

CONDUCTIMETRO		Jenway 3540	CODIGO	PML-038
Fecha	Patrón 100 $\mu\text{S}/\text{cm}$ (5 $\pm$)	Patrón 1413 $\mu\text{S}/\text{cm}$ (5 $\pm$)	Patrón 12900 $\mu\text{S}/\text{cm}$ (5 $\pm$)	Responsable
07.10.2015	101.1	1421	12 820	E.p
09.10.2015	100.1	1413	12 890	E.p
23.10.2015	103.0	1413	12 892	E.p
26.10.2015	109.5	1417	12 900	E.p.
24.11.2015	100.1	1416	12 930	E.p.
29.12.2015	100.70	1416	12 910	E.p
21.01.2016	99.7	1412	12 900	E.p
24.02.2016	99.5	1415	12 920	E.p
31.03.2016	100.2	1417	12.920	E.p
19.04.2016	101.7	1413	12.890	E.p
19.05.2016	100.8	1426	12.92	E.p
21.06.2016	100.9	1415	12.95	E.p
15.07.2016	100.1	1415	12.96	E.p
10.08.2016	100.2	1413	12.91	E.p
29.09.2016	100.3	1413	12.92	E.p
13.10.2016	100.1	1414	12.90	E.p
15.11.2016	98.8	1412	12.91	E.p
12.12.2016	100.5	1413	12.92	E.p